



# The Bhagirathi Co-operative Milk Producers' Union Limited

(A Govt. of West Bengal Project)

BU/PUR/QUOTATION/0533

May 23, 2023

## NOTICE INVITING QUOTATION

The Bhagirathi Milk Union is interested to purchase the following items from the accredited & authorized Suppliers/manufactures. Therefore, you are requested to quote your lowest rate. The specifications, quality and quantity of items to be purchased are given below:-

### (FORMAT-A)

Sl. No.	Particulars	Quantity	Unit Rate	GST	Other levies if any	Total Rate	Remarks
1.	<b>Hot Air Oven.</b> Technical Specification as per enclosed	One Unit					

I do hereby declare that rate quoted by me is not subject to any escalation during execution of this contract and I agree to execute the contract as per the rate quoted by me and payment terms enshrined in this NIQ.

### **Time line for Quotation:**

PARTICULAR	DATE & TIME
Date of Publication of NIQ	23.05.2023 at 11.00 am
Quotation submission start date & time	23.05.2023 from 11.05 am
Last date & time of submission of Quotation	30.05.2023 till 16.00 pm
Date & Time of opening of Quotation in the office of the Managing Director BCMPUL in the presence of willing bidders.	31.05.2023 at 15.00 pm

### **General Terms and Conditions:**

1. The bidders are required to go through the Terms & Condition carefully, understand the requirement and then quote the most competitive rate in sealed envelope addressed to the Managing Director, The Bhagirathi Cooperative Milk Producers' Union Limited, Feeder Dairy, Panchanantala, Berhampore, Murshidabad, 742101 (WB).
2. Quotation shall only be accepted in sealed envelope subscribing the Enquiry No. and Item Type. No quotations shall be accepted through mail. Incomplete quotations are liable to be rejected.
3. Price quoted by the bidder should be inclusive of all taxes levies packing, and forwarding F.O.R The Bhagirathi Cooperative Milk Producers' Union Limited and in adherence to the payment terms and other general terms and conditions enshrined in this NIQ. No escalation in price will be considered after submission of bids.
4. The bidders participating in the quotation must provide the following Mandatory Documents along with the price bids:
  - a. **Copy of NIQ duly signed and Sealed as a token of acceptance of all terms and conditions in the NIQ.**
  - b. **GST Registration Certificate**



# The Bhagirathi Co-operative Milk Producers' Union Limited

(A Govt. of West Bengal Project)

c. Documents in support of Experience Certificate and Credentials.

d. Rate as per format in FORMAT – A.

Failure to submit the above mandatory documents shall lead to rejection of the Quotation.

5. Materials must be delivered within 30 days after placement of Supply Order.
6. Mode of transportation for the supply of the goods to our site will be arranged by the bidder.
7. Materials dispatched from the bidder until or unless reaches the place of Delivery is the sole responsibility of the bidder. The Milk Union will not be responsible for any damage in transit, loading or unloading and no claim in this regards will be entertained.
8. Loading and Unloading of the goods will be the responsibility of the bidder and no claim regarding this will be entertained.
9. Complete documentation like principal warranty papers, operation manuals, technical specification and guarantee cards have to be furnished by the supplier.

## 10. Payment:

- 50% will be released on receipt of acknowledged copy of Challan/Invoice etc. of the whole unit in good condition at the specified site.
  - 40% payment will be released after Successful Installation, Commissioning & Trial-run of the whole unit duly certified by the official of Union.
  - 10% payment will be released after 6 months. This shall serve as performance guarantee.
11. The material received if found to be of substandard quality shall be rejected upon mutual inspection. In case material does not conform to the acceptability standard of the Milk Union the material has to be replaced by the party within 7 days.
  12. In case of any dispute between the Milk Union and the bidder the decision of the Milk Union shall be final and binding.
  13. The Milk Union reserves full right to accept or reject any quotation/all quotations or to cancel the entire process without assigning any reason thereof.

Sd/-

Managing Director  
The Bhagirathi Cooperative  
Milk Producers' Union Limited

Copy forwarded to:

1. Official Website ([www.bhagirathimilk.com](http://www.bhagirathimilk.com))
2. Office Notice Board

## **TECHNICAL SPECIFICATION**

HOT AIR OVEN	
Specification	Requirements
Capacity	30 ltr.
External Body	,Outer walls of Epoxy Powder coated GI sheets/textured SS
Internal Chamber	Stainless Steel 304 Grade



# The Bhagirathi Co-operative Milk Producers' Union Limited

(A Govt. of West Bengal Project)

Insulation	Mineral Wool/ Ceramic Wool
Door	<ul style="list-style-type: none"><li>• Inner: Stainless Steel 304 Grade</li><li>• Outer: Powder coated Mild Steel/ Stainless Steel 304 Grade</li><li>• Self-closing magnetic lock having door sealing material suitable to high temp</li></ul>
Adjustable Shelf	3 no. Perforated Stainless-Steel shelves (Removable) 304 Grade
Shelf Rest Pitch	30 mm
Temperature Range	<ul style="list-style-type: none"><li>• 40 °C to 200 °C</li><li>• Temperature should be thermostatically controlled</li><li>• Safety alarms</li></ul>
Least Count	0.1 °C
Temperature Accuracy	± 0.5 °C or better
Temperature Uniformity	±2 °C or better
Heating Element	Nichrome wire / Kanthal A1/SS tube/pipe heater
Time to attain Maximum Temperature	Approximately 90 minutes
Control Panel	Microprocessor based PID controller
Preset Timer	<ul style="list-style-type: none"><li>• With buzzer</li><li>• Digital display of time</li><li>• Least count- 1 minute</li></ul>
Circulation Method	Blower
Power Source	220-240 V, Single phase
Exhaust Port	30 mm ID on opposite side walls
Safety Device	<ul style="list-style-type: none"><li>• Self-diagnosis function including overshoot/undershoot of temperature and over current protection</li></ul>
Operating manuals,	Should provide:
service manuals, other manuals	<ul style="list-style-type: none"><li>• User, technical and maintenance manuals in English language</li><li>• List of equipment and procedures required for local calibration and routine maintenance</li><li>• Service and operation manuals to be provided</li></ul>
Recommendations or Warnings	Any warning signs would be adequately displayed
Warranty	Warranty for 2 years after satisfactory installation and working excluding consumable parts and accessories.
Training	The supplier will have to carry out successful Installation at the laboratory premises
List of Spares and Accessories	List of all spares and accessories (including minor) with part numbers and price, required for maintenance and repairs in future after guarantee/warranty period should be attached